



Clinisupplies

Clinisupplies(Nanjing) Co.,Ltd

No. 86-8, Shuanggao Road, Gaochun District
Nanjing, 211316 Jiangsu, P.R. China

Email: info@clinisupplies.co.uk
www.clinisupplies.co.uk

URGENT: PRODUCT RECALL NOTICE
(AKUT: PRODUKTÅTERKALLELSEMEDDELANDE)
Aquaflush Transanal Irrigation System

Date: 12 June 2026

FSCA Ref: CSN-FSCA-2026-1 | MMA Ref: FSCA054-2026

Tillverkare:

Clinisupplies (Nanjing) Co., Ltd.

No. 86-8 Shuanggao Road, Gaochun District, Nanjing 211316, Jiangsu, Kina

Produkt:

Aquaflush transanal irrigation system

Orsak till återkallelse:

Denna återkallelse beror på ett klassificeringsproblem enligt EU:s MDR (förordning (EU) 2017/745). Tillsatsen av en smörjgel påverkar produktens klassificering, vilket innebär att produkten inte helt uppfyller gällande regulatoriska krav.

Viktigt:

Det finns inga identifierade fel, funktionsstörningar eller några säkerhets- eller prestandaproblem med produkten. Eftersom produkten inte uppfyller de regulatoriska kraven får den dock inte användas.

Åtgärder som krävs:

- Sluta använda produkten
 - Returnera till distributören ELLER kassera enligt leverantörens anvisningar
- För support, vänligen kontakta din distributör.

Med vänliga hälsningar,

Ben Bain

Verkställande direktör, Clinisupplies (Nanjing) Co., Ltd.